

Weekly DI TESTING

COND= Conductivity
RES = Resistivity
TDS = Total Dissolved Solids
ORP_PPM = Free Chlorine

Record #	Measurement	Value	Unit	Date	Time
68	TDS	0	ppm	6/6/11	14:37:34
69	ORP_PPM	0.1	ppm	6/6/11	14:37:42
70	pH	7.01		6/6/11	14:37:57
71	COND	0.1	uS	6/13/11	12:39:04
72	RES	17.6	MOhm	6/13/11	12:40:19
73	TDS	0	ppm	6/13/11	12:40:25
74	ORP_PPM	0.1	ppm	6/13/11	12:40:41
75	pH	5.98		6/13/11	12:41:06
76	COND	0.09	uS	6/20/11	10:33:01
77	RES	17.8	MOhm	6/20/11	10:33:36
78	TDS	0	ppm	6/20/11	10:33:52
79	ORP_PPM	0.1	ppm	6/20/11	10:34:13
80	pH	6.15		6/20/11	10:34:18
81	COND	0.07	uS	6/27/11	12:45:25
82	RES	17.2	MOhm	6/27/11	12:45:35
83	TDS	0	ppm	6/27/11	12:45:42
84	ORP_PPM	0.1	ppm	6/27/11	12:46:00
85	pH	6.89		6/27/11	12:46:09
86	COND	0.09	uS	7/5/11	8:02:57
87	RES	18	MOhm	7/5/11	8:03:53
88	TDS	0	ppm	7/5/11	8:04:06
89	ORP_PPM	0.1	ppm	7/5/11	8:04:18
90	pH	6.66		7/5/11	8:04:24
91	COND	0.07	uS	7/12/11	10:53:13
92	RES	18.2	MOhm	7/12/11	10:54:42
93	TDS	5.65	ppm	7/12/11	10:55:00
94	ORP_PPM	0.1	ppm	7/12/11	10:55:13
95	pH	6.04		7/12/11	10:55:26
96	COND	0.07	uS	7/18/11	13:07:16
97	RES	17.5	MOhm	7/18/11	13:07:30
98	TDS	6.42	ppm	7/18/11	13:07:52
99	ORP_PPM	0.1	ppm	7/18/11	13:08:11
100	pH	7.18		7/18/11	13:08:19
101	COND	0.05	uS	7/25/11	13:16:14
102	RES	18.2	MOhm	7/25/11	13:16:18
103	TDS	0	ppm	7/25/11	13:16:22
104	ORP_PPM	0.1	ppm	7/25/11	13:16:25
105	pH	8.28		7/25/11	13:16:35
106	COND	0.06	uS	8/1/11	10:21:05
107	RES	17.7	MOhm	8/1/11	10:21:16
108	TDS	0	ppm	8/1/11	10:21:22

109	ORP_PPM	0.1	ppm	8/1/11	10:21:41
110	pH	7.35		8/1/11	10:21:52